

**IROAN STATEMENT OF WORK FOR THE
PULSE GENERATOR TEST SET
NSN 6625-01-161-0612 ID# 09047B**

1.0 SCOPE. This Statement of Work (SOW) establishes and sets forth tasks and identifies the work efforts that shall be performed by the Contractor (for purposes of this SOW, Contractor is defined as the commercial or government entity performing the IROAN) in the IROAN effort of the Pulse Generator Test Set. This document contains requirements to restore the Pulse Generator Test Set to Condition Code "A." Condition Code "A" is defined as "serviceable/issuable without qualification, new, used, repaired or reconditioned materiel which is serviceable and issuable to all customers without limitation or restriction, including materiel with more than six months shelf-life remaining."

1.1 Background. IROAN is defined as "That maintenance technique which determines the minimum repairs necessary to restore equipment components or assemblies to prescribed maintenance serviceability standards by utilizing all available diagnostic equipment and test procedures in order to minimize disassembly and parts replacement."

2.0 APPLICABLE DOCUMENTS. The following documents form a part of this SOW to the extent specified. Unless otherwise specified, the issues of these documents are those listed in the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto which is in effect on the date of solicitation. In the event of conflict between the documents referenced herein and the contents of the SOW, the contents of this SOW shall be the superseding requirement.

2.1 Military Standards

MIL-STD-129	DoD Standard Practice for Military Marking
MIL-STD-2073-1D	DoD Standard Practice for Military Packaging

2.2 Other Government Documents and Publications.

Military Handbook (For Guidance)

MIL-HDBK-61	Configuration Management Guidance
-------------	-----------------------------------

2.3 Industry Standards.

Mfr. 08116-90002	Operator's Manual
ANSI/ISO/ASQC Q9003-1994	Quality Systems - Model for Quality Assurance in Final Inspection and Test

Industry Standards (For Guidance)

ANSI/EIA-649

National Consensus Standard for Configuration Management

Copies of Military Standards and Specifications are available from the DOD Single Stock Point, Document Automation and Production Service, Building 4/D, 700 Robbins Avenue, Philadelphia, PA 19111-5094, Telephone (215) 697-2179 or DSN 442-2179, or <http://www.dodssp.daps.mil>. Copies of other government publications required by contractors in connection with specific SOW requirements shall be obtained through the Contracting Officer: Commander, Attn: Contracting Officer (Code 891), Marine Corps Logistics Bases, 814 Radford Blvd., Albany, GA 31704-1128, commercial telephone number (229) 639-6761 or DSN 567-6761. Copies of engineering drawings shall be obtained from: Life Cycle Management Center, Attn: (Code 851-3), 814 Radford Blvd Suite 20320, Albany, GA 31704-0320, commercial telephone number (229) 639-6410 or DSN 567-6410.

3.0 REQUIREMENTS

3.1 The Contractor shall repair, test and calibrate the Pulse Generator Test Set in accordance with Mfr. 08116-90002, Operator's Manual.

3.2 Quality acceptance and test procedures shall be in accordance with Mfr. 08116-90002, Operator's Manual.

3.3 Packaging, Handling, Storage and Transportation (PHS&T)

a. The Contractor shall be responsible for preservation and packaging of items being repaired under the terms of this statement of work. Items scheduled for long-term storage or shipment to overseas destinations shall be to level "A" requirements in accordance with MIL-STD-2073-1D, Appendix A, Table A.VI., Electronic Equipment. Items scheduled for domestic shipment, immediate use or short-term storage shall be to level "B" requirements.

b. Marking for shipment and storage shall be in accordance with MIL-STD-129.

c. The Marine Corps will provide the Contractor with the shipping address(es) for delivery of the repaired equipment. The Contractor shall be responsible for arranging for shipment to the pre-designated site(s). The Marine Corps will be responsible for transportation costs associated with shipping the subject equipment to and from the Contractor.

3.4 Quality Assurance Provisions. The contractor shall provide and maintain a Quality System that, as a minimum, adheres to the requirements of ANSI/ISO/ASQC Q9003-1994, Quality Systems - Model for Quality Assurance in Final Inspection and Test.

3.5 Configuration Control. The contractor shall apply configuration control procedures to established configuration items. The contractor shall not implement configuration changes to an

item's documented performance or design characteristics without prior written authorization. If it is necessary to temporarily depart from the authorized configuration, the contractor shall prepare and submit a Request for Deviation. MIL-HDBK-61 and ANSI/EIA-649 provide guidance for preparing this configuration control document.

(1 Data Item)

Form Approved
OMB No. 1704-0188

The Public reporting burden for this collection of information is authorized to average 110 hours per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302 and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503. Please DO NOT RETURN your form to either of these addresses. Send completed form to the Government issuing Contract Officer for the contract/PR No. listed in block E.

A. CONTRACT LINE ITEM NO.		B. EXHIBIT	C. CATEGORY: TDP _____ TM _____ Other <u>XXX</u>				
D. SYSTEM/ITEM Function Generator Test Set		E. CONTRACT/PR No.	F. CONTRACTOR				
1. DATA ITEM No. A001	2. TITLE OF DATA ITEM Request for Deviation (RFD)		3. SUBTITLE Configuration Management				
4. AUTHORITY (<i>Data Acquisition Document No.</i>) DI-CMAN-80640C		5. CONTRACT REFERENCE Paragraph 3.5		6. REQUIRING OFFICE MARCORLOGBASES (851)			
7. DD 250 REQ. I.T	9. DIST STATEMENT REQUIRED A	10. FREQUENCY AS REQ	12. DATE OF FIRST SUBMISSION See Blk 16	14. DISTRIBUTION			
8. APP CODE		11. AS OF DATE	13. DATE OF SUBSEQUENT SUBMISSION	a. ADDRESSEE			
				b. COPIES			
				Draft			
				FINAL			
				Reg			
				Repro			
16. REMARKS Blk 4: Contractor format is authorized and shall be submitted in .doc and .pdf format. Blk 10 & 12: RFDs will be reviewed and disposition determined within 30 working days upon receipt by the government. RFDs shall be transmitted via e-mail to the following address: <u>nbmamcomconfigmgmnt@matcom.usmc.mil</u> Distribution Statement A: Approved for public release, distribution is unlimited.				MCLBA (843-1)	0	0	0
				MCLBA (851-2)	0	1	0
				15. TOTAL			

17. PRICE GROUP

18. ESTIMATED
TOTAL PRICE

G. PREPARED BY:

H. DATE

I APPROVED BY:

DATE

DD FORM 1423-1, AUG 96 (EG)

PREVIOUS EDITION ON MAY BE USED

Page 1 of 1 Pages

Page 1 of 1 Pages
Designed using Perform Pro, WHS/D/or, Aug 96